




OXLIFE LIBERTY®

OXLIFE LIBERTY2®

User Manual



PORTABLE OXYGEN CONCENTRATOR

301-0005V Oxlife Liberty® and 301-0005VP Oxlife Liberty2®

USER: Read this manual before operating the device and save this manual for future reference.

HEALTH-CARE PROVIDER: This manual must be provided to the end user.



PROUDLY DESIGNED,
ENGINEERED AND
ASSEMBLED IN THE U.S.A.

Table of Contents

- Section 1: Introduction 5**
 - Intended Purpose..... 6
 - Intended Users 6
 - Intended Patient Population 6
 - Indications 6
 - Contraindications 6
 - Device and Accessory Symbols Reference 7
 - User Interface Symbols..... 8
 - Specifications 9
 - DNA Technology® 11
 - Applied Parts 11
- Section 2: Safety Guidelines..... 12**
 - Device Safety Guidelines 12
 - Patient Safety Guidelines 14
 - Battery and Power Supply Safety Guidelines 16
- Section 3: Product Description 17**
 - Device and Accessories 17
 - Feature Identification 18
 - Understanding the Control Panel 19
- Section 4: Operating Instructions 20**
 - Before Operating..... 20
 - Locating Your Device 21
 - Device Settings 22
 - Installing and Removing the Battery 25
 - Typical Battery Operation Times..... 25
 - Battery Time Management..... 26
 - AC Power Supply 27
 - DC Power Cord..... 29
 - DC Power Cord User Information Guide 31
 - DC Power Cord Troubleshooting Guide..... 32
 - Carry Bag..... 33
 - Accessory Bag 33

Cannula Use	34
Accessing the Menu Screen	35
Section 5: User Alerts and Alarms	37
Alarm and Alert Screen Descriptions.....	38
Alarm System Test.....	43
Section 6: Maintenance and Cleaning	44
Maintenance	44
Pre-Use Functional Check	44
Device Care and Cleaning.....	45
Section 7: Traveling	48
Entering and Exiting Airplane Mode	49
Section 8: Standards Compliance	50
Classification	50

Section 1: Introduction

PLEASE READ THIS USER MANUAL CAREFULLY BEFORE USING THE DEVICE. BE AWARE OF ALL WARNINGS AND SAFETY INFORMATION. ONLY USE ACCESSORIES APPROVED BY O₂ CONCEPTS® AND REFERENCED WITHIN THIS MANUAL.

IF YOU DO NOT FULLY UNDERSTAND ALL THE WARNINGS, SAFETY PRECAUTIONS, AND OPERATING INSTRUCTIONS, CONTACT YOUR AUTHORIZED DEALER OR HEALTHCARE PROVIDER FOR TECHNICAL SUPPORT.

Please contact your authorized dealer or healthcare provider if your Oxlife Liberty® requires service.

Please call O₂ Concepts® Technical Support to report any unexpected operation or events associated with the device at 1-877-867-4008 EXT 367 or contact the EU Rep or Authorized Representative listed on the rear cover of this user manual, or your country's relevant government agency, such as the competent authority.

Additional information about O₂ Concepts®, the Oxlife Liberty®, and the device limited warranty statement, can be found on our website at www.o2-concepts.com.

CAUTION: U.S. FEDERAL LAW RESTRICTS THESE DEVICES TO SALE BY OR ON THE ORDER OF A PHYSICIAN.

Intended Purpose

The Oxlife Liberty® is intended to administer supplemental oxygen delivered nasally to pediatric and adult patients. It is not intended for life support.

Intended Users

The Oxlife Liberty® is intended to be used by patients in a manner according to the intended purpose and operating guidelines. These devices are designed to be operated by patients who have received appropriate supervision and training, as needed, from healthcare professionals to ensure safe and effective use.

Intended Patient Population

The patient population includes pediatric and adult patients requiring supplemental oxygen as prescribed by a healthcare provider in the treatment of disorders aligning with the indications and contraindications of the device.

Indications

The Oxlife Liberty® is indicated on a prescription basis for the administration of supplemental oxygen used in the treatment of hypoxemia and other related conditions in patients with Chronic Obstructive Pulmonary Disease (COPD), Interstitial Lung Disease (ILD), Congestive Heart Failure (CHF), or other disease conditions where the administration of supplemental oxygen may be appropriate. It is not intended for life support. It is a small, portable device capable of continuous use in the home, institutional, and vehicle/mobile environments.

Contraindications

- ▲ Under certain circumstances, the use of non-prescription oxygen therapy can be hazardous. This device should only be used when prescribed by a healthcare provider.
- ▲ Not for use in the presence of aerosol sprays or flammable anesthetics.
- ▲ Additional monitoring may be required, as appropriate, for use by geriatric, pediatric, or any other patient unable to hear, see or interpret alarms, or communicate discomfort.
- ▲ The Oxlife Liberty® is not appropriate for any patient who would experience adverse health consequences as a result of a temporary interruption in oxygen therapy.
- ▲ The availability of an alternate source of supplemental oxygen is recommended in the case of a power interruption or a mechanical failure of the device. Consult your healthcare provider for their recommendation.
- ▲ It is the patient's responsibility to arrange for an alternate oxygen supply when traveling. O2 Concepts® assumes no liability for persons choosing not to adhere to manufacturer recommendations.
- ▲ Use of the device by patients with tracheotomy, under the recommendation of a qualified clinician, will require the use of appropriate clinical apparatus or attachments.

Device and Accessory Symbols Reference

Icon	Meaning	Icon	Meaning
	Caution represents the possibility of damage to the equipment		Warning represents the possibility of harm to the operator or patient
	No smoking while using or near device		Not suitable for use in the presence of a flammable anesthetic mixture
	No open flames		Do not incinerate battery
	Indoor Use Only*		Use no grease or oils
	Do not disassemble		Do not dispose of in household waste
	Refer to instructions		Keep dry in transport and storage
	Electronic instructions for use available		Class II equipment
	Type BF equipment		Prescription only
	Recycle battery		Manufacturer
IP22	Drip proof equipment		DC Power (Direct current)
	AC Power (Alternating current)		

Table 1 *Refers to AC power supply and DC power cord, not the Oxlife Liberty® device.

User Interface Symbols

Icon	Meaning	Icon	Meaning
	Power button		Cell connected
	Menu button		GPS/GNSS signal
	Increase flow setting Menu scroll		RDM connected (Tech Support Only)
	Decrease flow setting Menu scroll		Provider Information
	Battery charge status		Alarm
	AC/DC battery charging indicator		Breath detection
	Cell signal strength		Device status
	Exit		AC/DC power
	Airplane mode		Contains DNA Technology®
	Device Language		

Specifications

Dimensions	
Device Dimensions	H: 10 in (25.4 cm) W: 9 in (22.9 cm) D: 3.5 in (8.9 cm)
Weights	
Device Weight	6.36 lbs. (2.88 kg)
Battery	1.14 lbs. (0.52 kg)
AC Power Supply	1.60 lbs. (0.72 kg)
DC Power Cord	0.39 lbs. (0.17 kg)
Modes of Operation	
Continuous Mode	301-0005V Five Settings, 0.5 to 1.5 in 0.25 LPM increments.
	301-0005VP Seven settings, 0.5 to 2.0 in 0.25 LPM increments.
	Maximum deviation in flow rate is $\pm 10\%$. Flow maintained with outlet pressure ranging from 0 to 5.0 psig (0-34kPa)
Pulse Mode	301-0005V Settings 1 to 9 (160ml to 1500mL/minute) Bolus volumes are in 8mL increments, from 8 to 75mL.
	301-0005VP Settings 1 to 10 (160ml to 1600ml/minute) Bolus volumes are in 8mL increments, from 8 to 80mL.
	Maximum deviation of volume per breath is $\pm 15\%$ across all environmental conditions.
Breath Sensitivity	The device will deliver a bolus based on a trigger pressure range of -0.08 to -0.35 cmH ₂ O.
Battery Specification	Rechargeable 14.4V Lithium-Ion Battery
External Power Supply Power Input	AC power: 100-240 VAC, 50/60 Hz, 2.0A DC power: 11.5-15 VDC; Recommended: 10A outlet at 12V.
Operating Altitude	0-13,123ft or (0-4000m)

Oxygen Purity	91% $\pm$ 4% (87 - 95%); Measured purity values are within $\pm$ 2% of actual values; There is no variation in purity within the operating altitude. This applies to the full range of environmental operating conditions.
Operating Temperature	50°F (10°C) to 104°F (40°C)
Operating Humidity	10% - 95% @ 82.4°F (28°C)
Operating Atmospheric Pressure	1060hPa to 630hPa
Operating Environment	Free of smoke, pollutants, and fumes.
Transport/Storage Temperature	-4°F (-20°C) to 140°F (60°C)
Transport/Storage Humidity	0 -95% non-condensing
Operating Time	24 hours per day when connected to an external AC or DC power source. This is a continuous operation device.
Cannula Specification	DO NOT use cannula tubing longer than 7ft (2m) when using Pulse Flow Mode. DO NOT use cannula tubing longer than 50ft (15m) when using Continuous Flow Mode.
Sound Level	The typical volume at Pulse Setting 2 will be ~ 44.4 dBA. The volume that the device will reach at maximum settings will be ~ 52.3 dBA.
Alarm Sound Level	All alarms triggered by the device will be 75 dBA or higher at 10cm from the unit.
Maximum Outlet Pressure	Maximum outlet pressure is 7.4 psig (51kPa)
Service Life	Device: 5yrs, Accessories: 1yr

DNA Technology®

This version of the Oxlife Liberty® contains DNA Technology® (Dynamic Network Analysis), which is available to allow your health care provider to better serve your needs. This technology is intended to assist your provider in determining if the device is operating properly.



This device is FAA approved for use aboard passenger aircraft (FAA Advisory Circular 91.21-1B), including radio frequency emission limits of (RTCA) Document (DO) 160, section 21, Category M. Device contains FCC ID 2BRZXUBX19KM01.

The cellular connection must be deactivated prior to flight on any commercial aircraft.

Instructions for deactivating the cellular connection are included within [Section 7: Traveling](#).

There is no known interference posed by medical equipment during specific investigations or treatments.

There are no known devices that will cause interference issues.

Applied Parts

The applied parts, or components that a patient will come in contact with during normal operation of the device, are listed below:

- › Device Enclosure
- › Control Panel
- › Oxygen Outlet Port
- › External Power Input

Section 2: Safety Guidelines

Device Safety Guidelines

- ❶ **AVOID EXPOSURE TO OPEN FLAMES OR CREATION OF ANY SPARK NEAR YOUR OXLIFE LIBERTY. THIS INCLUDES SPARKS FROM STATIC ELECTRICITY CREATED BY ANY TYPE OF FRICTION. PROTECT ELECTRICAL POWER CORDS FROM SHARP EDGES TO AVOID ELECTRICAL SHOCK AND SERIOUS PHYSICAL INJURY.**
- ❶ Locate the Oxlife Liberty® in a well-ventilated area to allow for adequate air intake. Avoid the intake of airborne pollutants, smoke, or fumes. Wind or strong drafts can adversely affect accurate delivery of oxygen therapy.
- ❶ Use only approved accessories as specified in this User Manual to ensure proper function and avoid the risk of fire and burns. Use of non-approved accessories may cause serious damage to the device and will void the warranty.
- ❶ Locate oxygen tubing and power cords away from hot surfaces and in a manner to prevent tripping hazards.
- ❶ **DO NOT** modify the device. Any modifications made to the device may impair performance or damage the equipment and will void the warranty.
- ❶ **DO NOT** operate the device in an enclosed space.
- ❶ **DO NOT** cover the device or block the air inlet or the exhaust ports as it may impair device function.
- ❶ **DO NOT** drop or insert any objects or liquid into any opening.
- ❶ **DO NOT** leave your Oxlife Liberty® or batteries in your vehicle or trunk. Extreme heat or cold may damage your device and/or batteries.
- ❶ When shipping the Oxlife Liberty® and accessories, the batteries must be wrapped separately to avoid accidental battery contact. **DO NOT** ship the Oxlife Liberty® with the batteries installed.
- ❶ Keep the device away from potential household pests to avoid infestation that will impair device performance and will void the warranty.
- ❶ This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:
- ❶ This device may not cause harmful interference, and
- ❶ This device must accept any interference received, including interference that may cause undesired operation.
- ❶ Portable RF communications equipment (including peripherals such as antenna, cables, and external antennas) should be used no closer than 12 inches (30cm) to any part of the Oxlife Liberty® including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.
- ❶ Use of accessories, transducers, or cables other than those specified by O₂ Concepts® could result in increased electromagnetic emissions or decreased electromagnetic immunity of the device and result in improper operation.

- ❗ Use of the device adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, the device as well as other stacked equipment should be observed to verify that they are operating normally.
- ❗ The device is suitable to use in a CISPR 11 Group 1 Class B environment for radiated emissions.

Patient Safety Guidelines

- ▲ **PRESCRIBED MODE AND FLOW SETTINGS SHOULD ONLY BE ADJUSTED UNDER THE ADVICE OF A HEALTHCARE PROVIDER.**
- ▲ If you feel discomfort or are experiencing a medical emergency while undergoing oxygen therapy, seek medical assistance immediately to avoid harm.
- ▲ To ensure you are receiving the therapeutic amount of oxygen according to your medical condition, the Oxlife Liberty® must be used only after one or more settings have been individually determined or prescribed by a healthcare provider for you at your specific activity levels. The Oxlife Liberty® must be used with the specific combination of parts and accessories that were used while your settings were determined.
- ▲ **KEEP ALL MATCHES, LIT CIGARETTES, CANDLES, OR OTHER SOURCES OF IGNITION AT LEAST 10 FEET (3M) FROM THE DEVICE. THIS DEVICE PRODUCES ENRICHED OXYGEN GAS WHICH ACCELERATES COMBUSTION.**
- ▲ **DO NOT SMOKE WHILE USING THIS DEVICE. SMOKING DURING OXYGEN THERAPY IS DANGEROUS** and is likely to result in facial burns or death. **DO NOT** allow smoking within the same room where the oxygen concentrator or any oxygen carrying accessories are located. If you intend to smoke, you must always turn the oxygen concentrator off, remove the cannula and leave the room where the oxygen concentrator is located. If you are unable to leave the room, you must turn the oxygen concentrator off and remove the cannula before smoking.
- ▲ Use only water-based lotions or salves that are oxygen compatible before and during oxygen therapy. **DO NOT** use oil, grease, or petroleum-based products on or near the device to prevent accidental ignition.
- ▲ **DO NOT** lubricate fittings, connections, tubing or other accessories to avoid the risk of fire and burns.
- ▲ **DO NOT** leave the nasal cannula or mask on bed coverings or chair cushions. If the oxygen concentrator is turned on, but not in use, the oxygen will make the materials flammable. Turn the oxygen concentrator off when not in use to prevent oxygen enrichment.
- ▲ **DO NOT** operate the device on wet surfaces or in standing water, and do not submerge or expose to water or precipitation to avoid serious injury or damage to the device. Do not contact the device when wet to avoid serious injury or the chance of shock. If the Oxlife Liberty® has been dropped, damaged, or exposed to water please contact your authorized dealer or healthcare provider for inspection and possible service of the device.
- ▲ **DO NOT** use the device with an extension cord.
- ▲ **DO NOT** use the device with a damaged power cord or plug to avoid injury.
- ▲ Outdoor use must be conducted using battery power.
- ▲ Replace cannula as recommended by the cannula manufacturer. Cannulas may become contaminated with body fluids or expired gases during use.
- ▲ Monitor children and pets when in proximity to the device. Keep cannula tubing away from children and pets to avoid danger of choking or strangulation.

- ▲ Use of this device outside of the stated altitude, temperature or humidity ranges is expected to adversely affect the flowrate and the percentage of oxygen and, consequently, the quality of oxygen therapy.

Battery and Power Supply Safety Guidelines

- ❶ Use of non-approved battery or power supply may cause serious damage to the device and will void the warranty.
- ❷ Remove the battery from the device if it will not be used for an extended period of time to enhance battery life. Store the battery in a cool dry place when not in use to enhance battery life. Lithium-ion batteries can malfunction, especially when exposed to heat. Be sure to follow all safety recommendations provided in this user manual.
- ❸ **DO NOT** drop the battery or expose it to mechanical shock. If the battery is damaged, discontinue use and dispose of it properly to avoid personal injury.
- ❹ **DO NOT** expose the battery to water or other liquids. If the battery gets wet, discontinue use and dispose of it properly to avoid personal injury.
- ❺ **DO NOT** expose the battery to excessive heat or cold outside of specifications as it may affect performance.
- ❻ U.S. Department of Transportation (DOT) and United Nations (UN) regulations require that the battery be removed from the device when checked as luggage on international flights.
- ❼ The battery may explode if exposed to or disposed of in a fire. Keep the battery away from children to prevent injury.
- ❽ Locate the external power supply in the open air to prevent overheating when in use.
- ❾ **DO NOT** drive over, drag or place objects on the power cords to avoid damage.
- ❿ Use of a frayed or damaged AC or DC power supply cable may compromise basic safety.
- ⓫ **DO NOT** attempt to disassemble the battery or power supply. Doing so will void the warranty and may cause personal injury.
- ⓬ **DO NOT** use the battery or power supply for anything other than its intended purpose. Doing so may damage the device or cause personal injury.
- ⓭ Use caution when handling the DC plug adapter. This plug may get hot with use. Ensure DC plug socket is clean of debris which may cause overheating.

Section 3: Product Description

Device and Accessories

The Oxlife Liberty® is a portable oxygen concentrator used on a prescriptive basis that enables patients requiring supplemental oxygen to be treated in a home, institutional, and vehicle/mobile environment.

The Oxlife Liberty® is capable of delivering 87%-95% pure oxygen to a patient through a standard single lumen nasal cannula in both continuous flow and pulse mode.

Pulse mode detects a patient's breath and then delivers a bolus of oxygen during inhalation. In Continuous mode the device delivers a continuous flow of oxygen. Refer to the Specifications and other sections of this manual for details on specific settings and flow rates.

The Oxlife Liberty® has three standard power options, which include a 100-240V AC power supply, a 10-15V DC power cord, and a rechargeable lithium-ion battery.

Certain models may come equipped with O2 Concepts® DNA Technology® to provide equipment performance, usage data, and GPS/GNSS coordinates via cellular connectivity.

- ❗ The use of certain humidifiers and administration accessories not specified for use with this oxygen concentrator may impair performance. Your healthcare provider or authorized dealer will recommend the proper accessories for your use.
- ❗ Your healthcare provider or authorized dealer will recommend the proper cannula for your use. Ensure that the selected cannula is compliant to ISO 80601-2-69 Medical Electrical Equipment, Particular requirement for basic safety and essential performance of oxygen concentrator equipment.
- ❗ The configuration of the equipment and accessories must be determined for each individual patient.


This manual applies to the following accessories:

- › Battery
- › AC Power Supply
- › DC Power Cord
- › Accessory Bag
- › Carry Bag
- › Cannula (not included)

This manual does not apply to the following accessory items sold separately:

- › Desktop Battery Charger


Feature Identification



- 1. Control Panel
- 2. LCD Display
- 3. External Power Input
- 4. Rechargeable Battery
- 5. Oxygen Outlet Port

- 6. Side Vents
- 7. Exhaust Ports
- 8. Device Serial Number
- 9. FAA Approval Statement

Understanding the Control Panel



- 10. Power Button
- 11. Mode Button
- 12. Battery Status/Battery Charging Indicator
- 13. Cell Signal Strength
- 14. Airplane Mode
- 15. Cell Connected
- 16. GPS/GNSS Signal
- 17. RDM Connected (Tech Support Only)
- 18. RDM Mode (Tech Support Only)

- 19. Alert/Alarm Indication
- 20. Breath Detection
- 21. Green Status Circle – Open circle during warm-up. A Filled circle indicates flow and purity are within specifications.
- 22. External Power Source
- 23. Increase/Decrease & Menu Scroll Buttons
- 24. Device Setting
- 25. Mode Setting

Section 4: Operating Instructions

Before Operating

Your authorized dealer or healthcare provider has demonstrated the proper operation of your Oxlife Liberty®. This manual provides product information and operating instructions and should be saved for future reference to help you safely operate your device. If you have any further questions, contact your authorized dealer or healthcare provider.

This equipment must be placed into service in accordance with the information provided in the accompanying documents.

DO NOT operate the Oxlife Liberty® without first reading the Safety Guidelines included in **SECTION 2** of this manual.

Please follow all operating instructions.

Before each use, ensure that side vents are clean and in place and that cannula and power cord connections are intact.

If you are relocating your Oxlife Liberty® from an extreme environment, allow the device to return to the specified operating temperature and humidity ranges before use.

Operating your device outside of specified ranges may damage your device, impact device performance and may void your warranty.

See Operating Temperature and Humidity Ranges listed in the Specification Table.

Fully charge battery before first use.

The essential performance of the Oxlife Liberty® is to produce 91% ±4% oxygen at all continuous settings and pulse flow settings.

Locating Your Device

Place the Oxlife Liberty® in a well-ventilated area free of smoke, fumes, pollutants, and away from direct sunlight. Avoid high humidity environments.

Ensure side vents and exhaust ports are not obstructed.

Proper placement and positioning of the device is critical to the effectiveness of the oxygen therapy.

The Oxlife Liberty® **MUST** be located so that alarms can be heard.

Position the oxygen supply tubing and power cords in a manner that prevents kinking, air flow obstructions, and tripping hazards.

DO NOT locate the Oxlife Liberty® near any flammable materials or cleaning product or in the direct path of any heat source such as a stove, heat register or a car heater. Keep the Oxlife Liberty® at least 10 feet (3m) from hot sparking objects or open flame.

Device Settings

Turn the device on by pressing the Power button. 

Press the Mode button  to select Continuous or Pulse Mode.

Press the Increase  or Decrease  buttons to select the desired flow rate.

Turn the device off by pressing the Power button. 



When the device is powered on it will enter a five (5) minute warm-up period to reach desired performance. The warm-up period is indicated on the display screen by a green ring that becomes a green circle once warm up is complete.

THE PROPER FLOW MODE AND FLOW RATE ARE PRESCRIBED BY YOUR HEALTHCARE PROVIDER. DO NOT CHANGE THESE SETTINGS WITHOUT CONSULTING YOUR HEALTHCARE PROVIDER.

PULSE MODE SHOULD ONLY BE USED UNDER THE DIRECTION OF YOUR HEALTHCARE PROVIDER. APPROPRIATE AND SAFE PULSE MODE SETTINGS MUST ACCOMMODATE THE INDIVIDUAL PATIENT'S LIFESTYLE INCLUDING REST, TRAVEL AND EXERCISE.

DEVICE OXYGEN DELIVERY SETTINGS SHOULD BE PERIODICALLY REASSESSED FOR THE EFFECTIVENESS OF THE OXYGEN THERAPY.

THE PATIENT IS THE INTENDED OPERATOR.

The O2 Concepts Oxlife Liberty® is provided on a prescription basis for the administration of supplemental oxygen. Supplemental oxygen may reduce breathlessness and improve your ability to maintain an active lifestyle.

Consult your healthcare provider if you are feeling unwell, which may indicate too much or too little oxygen. These side effects are not immediate or life threatening. To monitor oxygen levels, your healthcare provider may recommend the use of a pulse oximeter.



Continuous (LPM)

CONTINUOUS MODE

In Continuous Mode the device will deliver a continuous flow of oxygen through the nasal cannula. The oxygen is measured in Liters per Minute or LPM.



Pulse

PULSE MODE

In Pulse Mode the device will detect your breath and deliver a measured pulse of oxygen or bolus through the nasal cannula. The breath detect icon will flash on the control panel to indicate breathing is detected.

Pulse Mode Bolus Volumes (mL)

Breaths per Minute	1	2	3	4	5	6	7	8	9	10*
15	8.0	16.0	24.0	32.0	40.0	48.0	56.0	65.0	75.0	80.0
20	8.0	16.0	24.0	32.0	40.0	48.0	56.0	65.0	75.0	80.0
25	6.4	12.8	19.2	25.6	32.0	38.4	44.8	52.0	60.0	64.0
30	5.3	10.7	16.0	21.3	26.7	32.0	37.3	43.3	50.0	53.3
35	4.6	9.1	13.7	18.3	22.9	27.4	32.0	37.1	42.9	45.7
40	4.0	8.0	12.0	16.0	20.0	24.0	28.0	32.5	37.5	40.0

*Pulse Mode setting 10 available with the Liberty2 device.

Pulse Mode No Breath Detection

When the device has not detected a breath for 45 seconds, the screen will turn amber, sound a 3-beep sequence and the text “No Breath” will flash on the screen. If the device does not detect a breath after an additional 15 seconds, the device will beep once and switch to Continuous Mode.

No Breath Detection Recovery

While in this condition, the breath detection icon on the top of the screen will turn red. The device will check for inspiration for 20 seconds every 15 minutes. If a breath is detected, the device will return to the last Pulse Mode setting. If no breath is detected, the device will stay in Continuous Mode and check for inspiration every 15 minutes.


The device will return to the last Pulse Mode setting at the next power up or if the power source is changed.

- ▲ Shallow respiratory effort of the patient may not trigger a bolus.
- ▲ The pulse flow settings of the Oxlife Liberty® do not correspond with continuous flow settings.
- ▲ The settings of other models or brands of oxygen therapy equipment do not correspond with the settings of the Oxlife Liberty®.

Installing and Removing the Battery

To install battery, align battery with battery slot in device and slide until battery clicks into place.

To remove the battery, depress the button and slide the battery toward you.



- ❗ Turn off the device before removing the battery.
- ❗ **DO NOT** use the Oxlife Liberty® without a battery installed.

Typical Battery Operation Times

Continuous Mode Settings	Operating Time/Hours
0.5	2.25
0.75	2.25
1.0	1.75
1.25	1.50
1.5	1.25
1.75*	1.00
2.0*	0.90
Pulse Mode Settings	Operating Time/Hours
1	3.00
2	2.75
3	2.50
4	2.00
5	1.75
6	1.50
7	1.30
8	1.10
9	1.10
10*	1.00

Battery operating times are based on a new, fully charged battery and are $\pm 10\%$.

Battery operating time will degrade with battery age, number of charge cycles and operating environment.

*Continuous Mode settings 1.75 and 2.0 and Pulse Mode setting 10 are available with the Oxlife Liberty2 device.

Battery Time Management

The Oxlife Liberty® is equipped with a rechargeable lithium-ion battery that is not user serviceable.

The battery will degrade over time from standard use. The rate of such degradation is dependent upon the frequency of operation of the device under battery power as well as the frequency and length of associated battery charging cycles.

The manufacturer recommends that the user regularly monitor battery performance by viewing the battery charge level indicator on the battery and the device display. Batteries should be replaced at least every two years.

With the combined use of the battery, AC power supply, and DC power cord you may extend your time away from home. Using the AC and DC power cords whenever possible will improve battery life. The battery will charge in the device, when plugged into an external power source.

Battery Charge Status is displayed on the control panel. As the charge level increases on the battery, the battery icon will turn solid in proportionate levels. When the battery is fully charged (over 95%) the battery icon will appear solid.

● 100

● 75

● 50

✱ 25

The Oxlife Liberty® battery includes a battery charge status indicator located on the front of the battery. Simply press the "PUSH" button on the battery to display the remaining battery life in 25% increments.

○
PUSH


The typical time to recharge a fully discharged battery is under 2.5 hours with the device plugged into the AC power supply and powered off. Battery charge times while the device is running will vary depending on the selected setting.

To maximize battery life:

- › Store battery in a cool, dry place when not in use. Store battery with at least 50% capacity remaining.
- › If using multiple batteries, uniquely label each battery to ensure each battery is rotated equally.
- › Charge batteries every three (3) months when not in use.

There is no routine maintenance or service required for the Oxlife Liberty® rechargeable batteries other than the periodic replacement as described above.

AC Power Supply



- ❗ Use only the AC power supply (800-1083) provided with this device. Use of power cords not supplied by O₂ Concepts® may cause overheating or damage to the device and will void the warranty.
- ❗ The power supply is not water resistant.
- ❗ **DO NOT** place anything in the power supply port other than the supplied AC or DC power cords.
- ❗ Ensure the power supply is in a well-ventilated area. The power supply may become hot during operation. Allow the power supply to cool before handling.

The AC power supply consists of the following components:

- › Power supply with attached power supply cable to connect to the device.
- › AC power input cable

When powered on, the green LED on the power supply will be illuminated.

The AC power supply charges the battery using a 100-240V 50/60 Hz outlet (a typical wall outlet in your home). The AC power supply allows use of your Oxlife Liberty® while simultaneously charging the battery.

To use the AC power supply, connect the power cord to the AC power converter (brick). Connect the power supply to a wall outlet and then to the Oxlife Liberty®. The external power icon will be displayed on the control panel.

The AC power supply will charge the battery at all settings.

Recommendation for Use

Do not use electrical outlets controlled by a switch.


When changing power sources, wait for the control panel to display the new power icon and the device to beep before removing the original source.

The detachable power supply cord to the power supply is the means of isolating the unit from the power supply main.

If the AC power supply is removed from the device, wait ten (10) seconds before re- applying power.

When traveling internationally, a standard, applicable, international power plug adapter will be required.

DC Power Cord



The DC power cord allows use of your Oxlife Liberty® device from a vehicle's 12-volt DC outlet in all settings while the vehicle is running. A 12-volt, 10-amp outlet is recommended for use with the device. The external power source icon will be displayed on the control panel. DC power is not sufficient to rapidly charge the battery at all settings. Details on battery charging speeds are provided according to the table below.

Device Settings	Flow Mode	Battery Charging Speed
0.5 – 1.0	Continuous	Rapid Charge
1 – 4	Pulse	
1.25 – 1.5	Continuous	Trickle Charge
5 – 10	Pulse	
1.75 – 2.0	Continuous	No Charge

Each vehicle's DC power outlet varies in specifications and performance. Moreover, such performance may be adversely impacted by one or more applications simultaneously drawing DC power (e.g., cell phones, media players, other electrical systems in the vehicle). In addition, DC power may fluctuate more significantly in stop and go driving conditions.

If your vehicle's DC voltage output drops below device power requirements, the device will alert. In this condition, the device's performance should be monitored. A flow setting reduction may improve DC performance. If a vehicle's DC voltage output continues to decrease, the device will alarm and then shut down. Switch to battery power to continue using your device.

- ① Use only the DC power cord (800-1079) provided with this device. Use of power cords not supplied by O2 Concepts® may cause overheating or damage to the device and will void the warranty.

Recommendations for Use

When operating the Oxlife Liberty® in your vehicle, ensure that the device is securely stowed and will not get damaged during transport.

Ensure that side vents and exhaust ports are not blocked.

The battery may not charge at any setting if the engine is not running.

Use the DC outlet closest to the car battery.

DO NOT leave the device plugged into the vehicle when the engine is not running.

DO NOT use the Oxlife Liberty® with any power splitting devices.

DC Power Cord User Information Guide

Starting the device on DC power

1. **ALWAYS** have the vehicle's engine running **BEFORE** plugging in your Oxlife Liberty®.
 2. Plug DC power cord into the device **BEFORE** plugging into the vehicle's DC outlet.
 3. Plug DC power cord into the vehicle's DC outlet.
 4. Remove the DC power cord from the vehicle's DC outlet when the engine is not running.
-

Disconnecting the device from DC power

1. When the device is OFF, disconnect the DC power cord from the vehicle's DC outlet.
The device will beep once and switch to internal battery power.

DC Power Cord Troubleshooting Guide

Event	Solution
Battery Not Charging	• Reduce the device mode to the lowest setting on either Pulse or Continuous mode and wait for charging icon to appear. • Slowly increase settings to determine if charging can be achieved at prescribed setting. Each vehicle's DC power performance varies and may limit achieving the desired flow rate. • Refer to DC Power Cord pg.29 for further details on how charging occurs at different flow settings.
Depleted Battery Charging	• A fully depleted battery may not begin charging unless the device is powered off. The lightning bolt may not display initially in the battery icon until the battery has charged to a minimum level.
Device Alerting / Beeping Intermittently	• Occurs when the DC voltage supplied by the vehicle drops in and out of the minimal DC input threshold required for device use. • Using a lower flow setting* may improve DC outlet performance.
Low External Power Alarm	• To reduce potential device shut down on DC power, eliminating other sources of power draw can improve performance. Air conditioning is an example of an electrical system that draws power away from the DC outlet. • Using a lower flow setting* may improve DC outlet performance. • Check your vehicle's user manual or consult an auto technician to determine if your vehicle's DC power amperage and wattage are sufficient. The device requires 12 volts (10 amps) to power reliably at all settings. • You may require an inverter with at least a 300-watt capacity. This can be purchased through any auto service provider.

** Consult your healthcare provider for advice on changing settings.*

Carry Bag

The Oxlife Liberty® carry bag provides protection and allows you to easily take your device with you for your daily activities.

Always use your Oxlife Liberty® in the supplied carry bag.

Check that the adjustable shoulder strap is secure, and the battery access zipper is closed before each use.



The Oxlife Liberty® is inserted into the carry bag from the bottom zippered section of the carry bag.

Ensure that side vents are aligned with the openings in the carry bag and access zipper is closed.

The battery can be accessed via the zipper without removing the device from the carry bag.

The carry bag is designed with an external pocket to carry your Oxlife Liberty® accessories or an additional battery.


Accessory Bag

The Accessory Bag is designed to carry your Oxlife Liberty® accessories including the AC power supply, DC power cord and extra batteries.



Cannula Use

- ▲ Your healthcare provider or authorized dealer will recommend the proper cannula for your use.
- ▲ Ensure selected cannula is compliant to
- ▲ *ISO 80601-2-69 Medical Electrical Equipment, particular requirement for basic safety and essential performance of oxygen concentrator equipment.*
- ▲ Use of a cannula not specified for use with oxygen may impair the performance of your device.
- ▲ **DO NOT** use cannula tubing longer than 7 feet (2m) in Pulse Flow Mode.
- ▲ **DO NOT** use cannula tubing longer than 50 feet (15m) in Continuous Flow Mode.
- ▲ A single lumen nasal cannula rated at 6 liters per minute is required to ensure proper oxygen delivery.
- ▲ To ensure proper oxygen flow, confirm that the cannula is not pinched or blocked in any way.
- ▲ The cannula fitting may be tight.
- ▲ **DO NOT** use grease or oils to lubricate the oxygen outlet port.



Read and follow the instructions included with the cannula provided by your authorized dealer.

Regularly clean and replace your cannula as instructed by the manufacturer and authorized dealer or healthcare provider.

O₂ Concepts® recommends the following cannulas by Salter Labs:


- › For Pulse Mode Use: [7ft](#) (2m) cannula: Part #16SOFT-7, or equivalent.
- › For Continuous Mode Use: [50ft](#) (15m) cannula: Part #OTC-203-50, or equivalent.

Accessing the Menu Screen

The provider screen shows information about the device's status and run hours.

To access the Menu screen:

- › Press the Mode (M) and Minus (-) buttons simultaneously.
- › Press the Plus (+) button or Minus (-) button to scroll through the screen options.
- › To enter and exit a screen, press the Mode (M) button.
- › Select the icon to enter the provider screen.



Run Hours:

The run hours are the total life hours the device has run.

(Example: Run Hrs: 69, where the device has been run a total of 69 hours)

```
SN:0097640001 08/18/22 11:27:19
Ver:4.27/4/6 Run Hrs: 69
D:4/0:200 R:0/0:0 Sg:22
T B1:46 M:47 B:33 O2:32
B1:21mA 8% 920 5

VIN:23.8 23.7 AlarmNone
Cell:Connected Mode:[cont]
Flow: 1.8 Pur:91
O2:0% PR:0 PF:0.0 -4:00:00 1/1
```

Voltage In:

Voltage In is the voltage that the device is receiving from an external power source such as the AC power supply or DC power cord.

(Example: VIN: 23.8 where 23.8 is in volts)

```
SN:0097640001 08/18/22 11:27:19
Ver:4.27/4/6 Run Hrs: 69
D:4/0:200 R:0/0:0 Sg:22
T B1:46 M:47 B:33 O2:32
B1:21mA 8% 920 5

VIN:23.8 23.7 AlarmNone
Cell:Connected Mode:[cont]
Flow: 1.8 Pur:91
O2:0% PR:0 PF:0.0 -4:00:00 1/1
```

Cell Status and Signal Strength:

The cell status is listed in the eighth row.
(Example: Cell: Connected)

The cell signal strength is listed in the third row on a scale of 0 – 31.
(Example: Sg:22)

```
SN:0097640001 08/18/22 11:27:19
Ver:4.27/4/6 Run Hrs: 69
D:4/0:200 R:0/0:0 Sg:22
T B1:46 M:47 B:33 O2:32
B1:21mA 8% 920 5

VIN:23.8 23.7 AlarmNone
Cell:Connected Mode:[cont]
Flow: 1.8 Pur:91
O2:0% PR:0 PF:0.0 -4:00:00 1/1
```

Changing Device Language

Device language is available in English, French, German, Italian and Spanish.

To change device language:

- › Press the Mode (M) and Minus (-) buttons simultaneously to access the Menu screen.
- › Press the Plus (+) button to scroll to the language icon (🌐).
- › Press the Mode (M) button and the current language will be displayed.
- › Press the Plus (+) button to scroll through the language choices.
- › Press the Mode (M) button to select language and then the Plus (+) button to confirm.

Section 5: User Alerts and Alarms

The functionality of the alarm system is verified automatically by the device upon start up by the display flashing three (3) visual display colors and audible indicators. There will be delays that are greater than ten (10) seconds inherent to specific alarms, and these delays are explained in detail for each alarm in the following tables. To ensure that audible alarm notifications are heard, the maximum distance the user can be from the device is dependent on surrounding noise levels. When two alarm conditions occur at the same time, the higher priority condition will be displayed.


Indefinite Acknowledge (Alarm Mute)

In an alert or alarm condition, pressing the power button momentarily will put the device into an indefinite acknowledge state or alarm mute. In this state, the alarm will be silenced, and the screen will stop flashing. The device will stay in this state until powered off or until a higher-priority alarm is activated.

DO NOT IGNORE ALERTS OR ALARMS

Alarm and Alert Screen Descriptions

Alarms







The LCD screen will be RED in Alarm Mode


High Priority alarms are indicated by a red screen flashing at 2Hz, and a 10-beep sequence repeated after 2.5 seconds of silence.

The device will continue to sound the alarm until the power  button is pressed momentarily to silence the audio; the power button is held to power off the device or 15 seconds has expired at which time the device will power off.

All alarms are technical, high priority alarms.

Low Purity Alarm  Low Purity	When oxygen levels drop below 72% for 60 seconds, the text “Low Purity” will flash on the screen. Solutions Clean or replace side vents. Change to another source of oxygen and contact your authorized dealer.
Low Battery Alarm  Low Battery	When battery power is depleted, the pump shuts off, and the text “Low Battery” will flash on the screen. Solutions Plug into an external power source. Replace depleted battery with a charged battery.
Low External Power Alarm  Low External Power	When the voltage from an external power source falls below 10 volts, as measured by the device internally, the text “Low External Power” will flash on the screen. Solutions Ensure all charging connections are intact. Ensure the battery is installed. Refer to the DC user guide section of this manual. Change to another source of power and contact your authorized dealer.

Unauthorized Battery Alarm  Unauthorized Battery	If the device detects a battery other than an Oxlife Liberty[®] battery has been installed in the device, the battery icon will show an exclamation point within the battery outline. If no other power source is present, the text “Unauthorized Battery” will flash on the screen. Solutions Remove the unauthorized battery and replace it with an authorized Oxlife Liberty[®] battery.
Over Temperature Alarm  Over Temperature	The maximum operating temperature of the enclosure and pump is 70°C. If either of these components reach 70°C for 15 seconds, the text “Over Temperature” will flash on the screen. Solutions Relocate the device to improve airflow or to a cooler environment and allow the device to reach operating temperature. Utilize an alternate source of oxygen if necessary and contact your authorized dealer.
Invalid Motor Temperature Alarm  Invalid Motor Temperature	When the motor temperature is out of a valid range for 15 seconds, the text “Invalid Motor Temperature” will flash on the screen. Solutions Relocate the device to improve airflow or to a cooler environment and allow the device to reach operating temperature. Utilize an alternate source of oxygen if necessary and contact your authorized dealer.
Invalid Box Temperature Alarm  Invalid Box Temperature	When the internal box temperature is out of a valid range for 15 seconds, the text “Invalid Box Temperature” will flash on the screen. Solutions Relocate the device to improve airflow or to a cooler environment and allow the device to reach operating temperature. Utilize an alternate source of oxygen if necessary and contact your authorized dealer.

Invalid Flow Alarm 	When the pulse volume is out of a valid range by 15% for 60 seconds the text "Invalid Flow" will flash on the screen. Solutions Check the cannula connection. Repair or replace tubing. Clean or replace side vents. Move device to ensure adequate air flow. Ensure the device is oriented for operation, vertically with display visible or horizontally. Change to another source of oxygen and contact your authorized dealer.
No Flow Alarm 	If after 30 minutes the "No Flow" Alert condition has not been corrected, the device will initiate a No Flow Alarm and shut down after 15 seconds. If the No Flow condition is corrected, the device will return to normal function. Solutions Check the cannula connection. Repair or replace tubing. Clean or replace side vents. Move device to ensure adequate air flow. Change to another source of oxygen and contact your authorized dealer.

DO NOT IGNORE ALERTS OR ALARMS

Alerts




The LCD screen will be AMBER in Alert Mode

Medium Priority alarms are indicated by an amber screen and a 3-beep sequence repeated after 15 seconds of silence.

Low Priority alarms are indicated by an amber screen and a single beep every 30 seconds. Pushing the power button once silences the alert.

The Unauthorized Battery alert is a mechanical alert. All other alerts are technical. The Low Battery and Low Purity alerts will sound per the low priority alarm sequence. All other alerts employ the medium priority alarm sequence.

No Breath Alert



When the device has not detected a breath for 45 seconds the text "No Breath" will flash on the screen.

If the device does not detect a breath after an additional 15 seconds, the device will beep once and switch to Continuous Mode. While in this condition, the breath detection icon on the top of the screen will turn red.

The device will check for inspiration for 20 seconds every 15 minutes. If a breath is detected, the device will return to the last Pulse Mode setting. If no breath is detected, the device will stay in Continuous Mode and check for inspiration every 15 minutes.


The device will return to the last Pulse Mode setting at next power up or if the power source is changed.

Solutions

Check the cannula for kinks or obstructions.

Ensure patient is breathing through their nose.

Low Purity Alert






When oxygen levels drop below 85% for 60 seconds, the text "ALERT Low Purity" will flash on the screen.

Solutions

Clean or replace side vents.

Contact your authorized dealer.


Low Battery Alert  Low Battery	When calculated battery run time reaches 6 minutes remaining, the text "ALERT Low Battery" will flash on the screen. Solutions Connect to an external power source. Replace depleted battery with a charged battery.
Unauthorized Battery Alert  Unauthorized Battery	If the device detects that a battery other than an Oxlife Liberty[®] battery has been installed in the device, the battery icon will show an exclamation point within the battery outline. The device will continue to run if AC or DC power is present. The text "Unauthorized Battery" will flash on the screen. Solutions Remove the unauthorized battery and replace it with an authorized Oxlife Liberty[®] battery.
Low External Power Alert  Low External Power	When the voltage from an external power source falls to 10.5 volts or less, as measured by the device internally, for 3 seconds, the text "ALERT Low Ext Power" will flash on the screen. Solutions Ensure all charging connections are intact. Ensure the battery is installed. Refer to the DC user guide section of this manual. Change to another source of power and contact your authorized dealer.
NoFlow Alert  No Flow	When the flow of oxygen has stopped for 45 seconds, the text "No Flow" will flash on the screen for up to 30 minutes. If the No Flow condition is corrected, the device will return to normal function. If at the end of 30 minutes the No Flow condition has not been corrected, the device will initiate a No Flow Alarm. Solutions Check the cannula connection. Repair or replace tubing. Clean or replace side vents. Move device to ensure adequate air flow. Change to another source of oxygen and contact your authorized dealer.

DO NOT IGNORE ALERTS OR ALARMS

Alarm System Test

Use this test to verify that the alarm system is working properly. This test includes screen color, screen flashing sequence, and audible alarm sequence. Follow the steps in the procedure listed below.

- › Simultaneously press the Mode (M) button and Minus (-) button to enter the Menu screen.
- › Press the Plus button (+) or Minus button (-) to scroll to the Alarm Test icon.
- › Press the Mode button (M) to select Alarm Test.




The system will begin by testing the High Priority Alarm. The High Priority Alarm consists of a 10-beep sequence separated by 2.5 seconds of silence and the LCD screen flashes red. The alarm symbol will appear on the screen.

The device will automatically enter the Medium Priority Alarm test once the High Priority Alarm sequence has concluded. Pressing the Mode (M) button while the High Priority Alarm test is in progress will terminate the High Priority Alarm test and start the Medium Priority Alarm. The Medium Priority Alarm consists of a 3-beep sequence separated by 15 seconds of silence and the LCD screen will flash amber. The alarm symbol will appear on the screen.

The device will automatically enter the Low Priority Alarm test once the Medium Priority Alarm sequence has concluded. Pressing the Mode button (M) while the Medium Priority Alarm test is in progress will terminate the Medium Priority Alarm test and start the Low Priority Alarm. The Low Priority Alarm consists of a single, 1-beep and a solid amber LCD screen. The alarm symbol will appear on the screen.

At the conclusion of the test, the menu screen will appear. Use the Plus button (+) or Minus button (-) to toggle to the Home screen. Press the Mode button (M) to return to the home screen.



The Remote Diagnostics and Cell Info screens are for use solely by O₂ Concepts® or your provider's technical support.

Section 6: Maintenance and Cleaning

Maintenance

Routine cleaning of the side vents, and the device care and cleaning, as described below, are the only routine maintenance by the user necessary for the operation of the Oxlife Liberty®.

There is no routine maintenance or service to the Oxlife Liberty® lithium-ion rechargeable batteries other than periodic replacement as described in the section on Battery Time Management, page [26](#).

All other maintenance or service **MUST** be conducted by a qualified Oxlife Liberty® service technician. **DO NOT** attempt to disassemble or perform any maintenance on your device. Any such attempt will void the warranty.

Pre-Use Functional Check

A trained service technician can qualitatively verify that the device is ready for use by performing the following protocols:


- › Connect the nasal cannula to the cannula port on the device. Set the device to Continuous Mode. The technician should be able to hear the flow of oxygen by placing the outlet of the nasal cannula near his or her ear (do not direct flow straight at ear to avoid discomfort).
- › The technician should be able to feel the flow of oxygen by placing a finger roughly a half inch from the outlet of the nasal cannula.
- › Place the end of the nasal cannula in a half-full cup of water. Verify that there are bubbles from the output of the nasal cannula.

Device Care and Cleaning

Side Vent Cleaning

Routine cleaning of the side vents as described below and periodic cleaning of the exterior of the device is the only routine maintenance.

To keep your Oxlife Liberty® working properly, it is recommended you clean the side vents weekly. If the Oxlife Liberty® is used in a dusty environment, more frequent cleaning of the side vents may be required.



Follow these simple directions to clean the side vents:

1. Remove the side vent.
 2. Wash the vent by running under warm tap water using a mild detergent.
 3. Rinse thoroughly under warm, running water.
 4. Tap the side vent on a counter or hard surface several times to remove excess water.
 5. It is important to allow the vent to dry completely before reinserting into the device. Moisture from wet or damp side vents may damage your device.
- ❗ Never use your Oxlife Liberty® without side vents installed.
- ❗ We recommend that you keep extra side vents to use as a replacement while one is drying. Contact your authorized dealer for extra side vents.

Device and Battery Cleaning

- ▲ Unplug your Oxlife Liberty® and remove the battery before cleaning.
- ▲ Do not submerge your device or any accessories or allow water to enter the device.

Clean the exterior of your device and/or battery using a soft cloth dampened with a mild detergent and water and wipe dry. Clean the outside of the device monthly or more frequently, as needed.

- ❗ **DO NOT** spray or soak the case or front panel.
 - ❗ **DO NOT** spray or soak battery contacts.
 - ❗ **DO NOT** use alcohol, ethylene chloride, benzalkonium chloride or petroleum-based cleaners.
-

AC Power Supply and DC Power Cord Cleaning

- ▲ Unplug your Oxlife Liberty® and/or power supply before cleaning.

Clean the AC power supply and DC power cord using a soft cloth dampened with a mild detergent and water and wipe dry.

Carry Bag Cleaning

- ❗ Remove the device from the carry bag before cleaning.

Follow these simple directions to clean the Carry Bag:

1. The carry bag can be cleaned using a damp cloth or by washing in a washing machine alone, on a gentle cycle, using cold water and a mild detergent.
 2. Allow the carry bag to air dry thoroughly.
 3. Ensure the carry bag is completely dry before use.
-

Cannula Cleaning

Clean and replace the cannula as instructed by the manufacturer and your healthcare provider.

Disinfection

Between patients, the exterior of the device, carry bag, and battery should be disinfected.

Clean the item as instructed above.

Per manufacturer's instructions, disinfect device exterior, carry bag and battery using Oxycide (or similar product) with a soft cloth.

- ❗ **DO NOT** use alcohol, ethylene chloride, benzalkonium chloride or petroleum-based cleaners.
-

Device Disposal

This product may contain substances that could be harmful to the environment if inappropriately disposed of in landfills. Follow local governing ordinances and recycling plans regarding disposal of the device.

Battery Disposal

The battery is rechargeable and can be recycled. Always return batteries to your healthcare provider or an authorized dealer for proper disposal. You can also contact your local government offices for proper disposal instructions for a lithium-ion battery.

Section 7: Traveling

Before traveling, be sure to pack the following:

- › Accessory Bag
- › Carry Bag
- › AC Power Supply
- › DC Power Cord
- › Fully Charged Battery (and extra batteries if required)
- › User Manual

Also, bring contact information for your healthcare provider or authorized dealer.

When traveling internationally, an applicable standard international power plug adapter is required.

Please note that many air carriers require prior notification (usually at least 48 hours) for passengers traveling with portable oxygen concentrators.


In addition, carriers may also require advanced submission of a medical verification statement from your healthcare provider. You should confirm any notification and verification requirement with your specific carrier.

Not all air carriers provide an electrical outlet aboard the aircraft, so you should have sufficient batteries (or a confirmed alternative power source) to account for at least 150% of your total travel time, including, but not limited to: (A) commuting time to and from the airport; (B) transition time in and out of airports; (C) the duration of your flight; and (D) any unexpected delays. Please be aware that battery requirements may vary based upon your personal setting.

If traveling by train, bus or boat contact your carrier to inquire about power port/outlet availability.

Entering and Exiting Airplane Mode

The cellular connection must be deactivated prior to flight on any commercial aircraft.



To enter and exit Airplane Mode, follow the steps listed in the procedure below.

1. Simultaneously press the Mode (M) and Minus (-) buttons to enter the Menu screen.
2. Press (+) or (-) buttons to scroll to Airplane Mode icon.
3. Press Mode button (M) to enter Airplane Mode menu.
4. Press Plus button (+) to activate Airplane Mode.
5. Press Mode button (M) to select displayed mode.
6. Press Plus button (+) to confirm or Minus button (-) to cancel.
7. Press Plus button (+) or Minus button (-) to scroll to the Home icon.
8. Press Mode button (M) to return to the Home screen.

The device will exit Airplane Mode after 24 hours or the user manually re-enters Normal Mode.

To manually re-enter Normal Mode, follow instructions above.

Section 8: Standards Compliance

IEC 60601-1: 2005 Ed.3+A1; A2 Medical Electrical Equipment; Part 1: General Requirements for Basic Safety and Essential Performance

IEC 60601-1-2: 2020 Ed. 4.1; Medical Electrical Equipment, Part 1-2: General Requirement for Safety – Collateral Standard: Electromagnetic (EMC) Compatibility

IEC 60601-1-6:2010 Ed3+A1; A2 Medical Electrical Equipment – Part 1-6 General Requirements for Basic Safety and Essential Performance – Collateral Standard: Usability

IEC 60601-1-8:2006 Ed.2 + A1; A2 Medical Electrical Equipment – Part 1-8: General Requirements for Basic Safety and Essential Performance – Collateral Standard: General Requirements Tests and Guidance for Alarm Systems in Medical Electrical Equipment and Medical Electrical Systems

IEC 60601-1-11:2015 Ed.2+A1 Medical Electrical Equipment Part 1-11: General Requirements for Basic Safety and Essential Performance Collateral Standard – Requirements for Medical Electrical Equipment and Medical Electrical Systems Used in the Home Healthcare Environment

IEC 62304:2006 Ed1 + A1 Medical Device Software – Software Life Cycle Processes

ANSI/AAMI IEC 62366-1:2015 Ed.1+A1 Medical Devices – Application of Usability Engineering to Medical Devices

ISO 80601-2-67:2020 Ed.2 Medical Electrical Equipment – Part 2-67 Particular Requirements for Basic Safety and Essential Performance of Oxygen Conserving Equipment

ISO 80601-2-69:2020 Ed.2 Medical Electrical Equipment – Part 2-69 Particular Requirements for Basic Safety and Essential Performance of Oxygen Concentrator Equipment

RTCA, DO 160, Section 21, Category M; Emission of Radio Frequency Energy

ISO 13485:2016 Medical Devices; Quality Management Systems; Requirements for Regulatory Purposes

ISO 15223-1:2021/Amd.1:2025 Medical Devices – Symbols to be used with medical device labels, labelling and information to be supplied by the manufacturer – Part 1: General Requirements

Classification

The Oxlife Liberty® is classified as:

- › IEC Class II Internally Powered Equipment
- › Type BF Applied Part
- › IP22: Drip Proof Equipment (Protected against solid objects over 12 mm and direct sprays of water up to 15° of vertical per IEC 60529)
- › **NOT** suitable for use in the presence of flammable anesthetic mixture with air or with oxygen or nitrous oxide
- › Continuous Operation

OXLIFE LIBERTY®



O2 Concepts, LLC
199 Park Road Extension
Middlebury, CT 06762 USA
+1(877) 867-4008
www.o2-concepts.com



MED IVD HEALTHTECH SL
C/Alejandro Casona, 2 portal 6-3 b 2,
28702 San Sebastian de los Reyes,
Madrid, Spain
+34 93 609 62 69
Email: ecrep@mdxcro.com



2797



ENERGY SMART
TECHNOLOGY®

©2026 O₂ Concepts®, LLC, all rights reserved; O₂ Concepts®, Oxlife Liberty®, DNA Technology®, and Energy Smart Technology® including certain related logos, are registered trademarks of O₂ Concepts®, LLC and may not be reproduced without express written consent.

PN: 800-1075 Rev F – 2026-05-27